



Column Name: FLARE WP C18  
 Column Dimensions: 2.1 x 50 mm  
 Column ID: 16636.6-2 wide pore  
 HPLC System: Agilent 1290  
 Injection Volume: 1.0  $\mu$ l  
 Detection: UV at 280 nm  
 Flow Rate: 0.2 ml/min or 1.0 ml/min  
 Mobile Phase: A: 5 ml ACN, 95 ml H<sub>2</sub>O, 2 ml TFA; B: 95 ml ACN, 5 ml H<sub>2</sub>O, 2 ml TFA

Gradient: Time (mins) %A %B (0.2 ml/min)

0.00	100	0
50.00	0	100
50.01	100	0
44.00	100	0 end

Gradient: Time (mins) %A %B (1.0 ml/min)

0.00	100	0
10.00	0	100
10.01	100	0
11.00	100	0 end

Temperature: 60 °C

Analyte: Bovine serum albumin (MW= 66.5 kDa)

