



Column Name: FLARE WP C18

Column Dimensions: 4.6 x 33 mm

Column ID: 16636.2-3-1 wide pore

HPLC System: Agilent 1200

Injection Volume: 5.0 μ l

Detection: UV at 230 nm

Flow Rate: 1.2 ml/min

Mobile Phase: A: 5 ml ACN, 95 ml H₂O, 0.1 ml TFA; B: 80 ml ACN, 20 ml H₂O, 0.1 ml TFA

Gradient: Time (mins) %A %B

0.00 100 0

3.00 30 70

3.01 100 0

6.00 100 0 end

Analyte: Deoxyribonuclease I from bovine pancreas (MW= 39 kD)



DiamondAnalytics