



Column Name: FLARE WP C18  
 Column Dimensions: 4.6 x 100 mm  
 Column ID: 16636.1-3 wide pore  
 HPLC System: Agilent 1200  
 Injection Volume: 5.0  $\mu$ l  
 Detection: UV at 230 nm  
 Flow Rate: 1.0 ml/min  
 Mobile Phase: A: 5 ml ACN, 95 ml H<sub>2</sub>O, 0.1 ml TFA; B: 80 ml ACN, 20 ml H<sub>2</sub>O, 0.1 ml TFA  
 Gradient:

Time (mins)	%A	%B
0.00	100	0
14.00	60	40
14.01	100	0
22.00	100	0

end  
 Temperature: 35 °C  
 Analyte: Bovine hemoglobin (MW= 60.4 kD)