



Column Name: FLARE WP C18

Column Dimensions: 4.6 x 100 mm

Column ID: 16636.1-3 wide pore

HPLC System: Agilent 1200

Injection Volume: 5.0  $\mu$ l

Detection: UV at 230 nm

Flow Rate: 1.0 ml/min

Mobile Phase: A: 5 ml ACN, 95 ml H<sub>2</sub>O, 0.1 ml TFA; B: 80 ml ACN, 20 ml H<sub>2</sub>O, 0.1 ml TFA

Gradient: Time (mins)	%A	%B
0.00	100	0
12.00	60	40
12.01	100	0
20.00	100	0 end

Temperature: 35 °C

Analyte: Histone from calf thymus II-A (MW= 11-21 kD)



DiamondAnalytics