



Column Name: FLARE WP C18

Column Dimensions: 2.1 x 50 mm

HPLC System: Agilent 1200

Injection Volume: 2.0 μ l

Detection: UV at 280 nm

Flow Rate: 1.0 ml/min

Mobile Phase: A: x% TFA in H₂O, B: x% TFA in ACN

Gradient:

Time (mins)	%A	%B
0.00	100	0
5.00	0	100
5.01	100	0
7.00	100	0 end

Temperature: 60°C

Analyte: Lysozyme, chicken egg white



DiamondAnalytics