



Column Name: FLARE WP C18
 Column Dimensions: 4.6 x 33 mm
 Column ID: 16636.2-3-1 wide pore
 HPLC System: Agilent 1200
 Injection Volume: 5.0 μ l
 Detection: UV at 230 nm
 Flow Rate: 1.2 ml/min
 Mobile Phase: A: 5 ml ACN, 95 ml H₂O, 0.1 ml TFA; B: 80 ml ACN, 20 ml H₂O, 0.1 ml TFA
 Gradient: Time (mins) %A %B

Time (mins)	%A	%B
0.00	100	0
3.00	30	70
3.01	100	0
6.00	100	0 end

Analyte: Ribonuclease A from bovine pancreas (MW= 13.7 kD)