



Column Name: FLARE WP C18
Column Dimensions: 4.6 x 100 mm
Column ID: 16636.1-3 wide pore
HPLC System: Agilent 1200
Injection Volume: 5.0 μ l
Detection: UV at 230 nm
Flow Rate: 1.0 ml/min
Mobile Phase: A: 5 ml ACN, 95 ml H₂O, 0.1 ml TFA; B: 80 ml ACN, 20 ml H₂O, 0.1 ml TFA

Gradient: Time (mins)	%A	%B
0.00	100	0
12.00	60	40
12.01	100	0
20.00	100	0 end

Temperature: 55 °C

Analyte: Trypsin inhibitor from Glycine max (soybean) (MW= 20.1 kD)