



Column Name: FLARE WP C18

Column Dimensions: 4.6 x 100 mm

Column ID: 16636.1-3 wide pore

HPLC System: Agilent 1200

Injection Volume: 5.0 μ l

Detection: UV at 230 nm

Flow Rate: 1.0 ml/min

Mobile Phase: A: 5 ml ACN, 95 ml H₂O, 0.5 ml TFA; B: 80 ml ACN, 20 ml H₂O, 0.5 ml TFA

Gradient: Time (mins) %A %B

0.00 100 0

12.00 60 40

12.01 100 0

19.00 100 0 end

Temperature: 35 °C

Analyte: Whey from bovine milk (MW= 14 - 65 kD)



DiamondAnalytics